

Seated position

Q When using the FLOWTRON® system, does the seated position adversely affect venous return?

A No. We undertake testing of the *FLOWTRON* system with all our associated garments, in subjects with limbs elevated, as well as dependent and both positions demonstrate significant venous augmentation. However having the legs elevated does enhance venous return.

The answers to the commonly asked questions may inform and guide but they are not a substitute for holistic patient assessment and clinical judgment. If you cannot find an answer for which you are looking, please register your query at www.act-club.info